



STATE OF ARKANSAS
OFFICE OF STATE PROCUREMENT
 1509 West 7th Street, Room 300
 Little Rock, Arkansas 72201-4222

ADDENDUM 1

TO: Vendors Addressed
 FROM: Judy Shirley, Buyer
 DATE: March 30, 2017
 SUBJECT: SP-17-0108

The following change(s) to the above-referenced IFB have been made as designated below:

- ☒ Change of specification(s)
- ☐ Additional specification(s)
- ☐ Change of bid opening time and date
- ☐ Cancellation of bid
- ☐ Other

CHANGE OF SPECIFICATIONS

- Delete Section 3.2.A and replace with the following:
 - A. Transmitted Light Comparison Microscope Bridge
 - Comparison images can be viewed as 100% right, 100% left, divided, or overlapped in any ratio
 - Images **must** be erect and unreversed
 - Camera port included
 - **Must** have color and intensity balance between both microscopes without the use of additional filtering or adjustment by the examiner
 - Field Number 23

The specifications by virtue of this addendum become a permanent addition to the above referenced IFB. Failure to return this signed addendum may result in rejection of your bid.

If you have any questions please contact Judy Shirley at judy.shirley@dfa.arkansas.gov or (501) 324-9316.

Company: _____

Signature: _____

Date: _____